

**PYROMETRIC COMPANY**1312 South 96TH Street
Seattle, Washington 98108

(206) 762-8307 www.pyrometric.com Fax (206) 763-9459

CALIBRATION REPORT

Customer: WASHINGTON STATE PATROL P.O.# 00081650 Test No. 19577
Location: PROPERTY MANAGEMENT CONTROL Attn: GREG PARRIOTT Test Date: 11/09/04
Instrument: TOLEDO LOAD CELL C2P1-50K/RICE LAKE INDICATOR IQ PLUS 310 Recall Date: 11/09/06
Serial No. 94196/91441281 Range: 25,000 LBF. Req'd Accy: $\pm 1\%$ Indicated

As Found Condition:	In Tol. <u>X</u>	Out of Tol. _____	Repair _____
STANDARD	INSTRUMENT	CORRECTION	
0.0 LBF.	0 LBF.	0.0 LBF.	
2499.3	2500	-0.7	
4997.4	5000	-2.6	
7495.7	7500	-4.3	
9994.3	10000	-5.7	
12495.0	12500	-5.0	
14994.0	15000	-6.0	
17495.3	17500	-4.7	
19995.9	20000	-4.1	
22496.6	22500	-3.4	
25000.0	25000	0.0	

NOTE: READINGS OBTAINED FROM AVERAGE OF 2 RUNS

I Certify, Under Penalty of Perjury, Under the Laws Governing the State of Washington that the foregoing is True and Correct.

Condition at time of testing: Temperature: 68°F Relative Humidity: 46 %

Remarks: The above equipment meets or exceeds the requirements of the following specifications:

PER. CUSTOMER _____, _____, _____, _____, _____

Test procedure used: 4.3.2. Our calibration system complies with ANSI/NCSL Z540-1.
All test equipment used is traceable to standards of the National Institute of Standards and Technology.

Test Equipment	Serial No.	Test Date	Recall Date	Test No.
REVERE LOAD CELL	707481	04/15/04	04/15/05	17555
INTERFACE LOAD CELL	117123A	04/16/04	04/16/05	17555
_____	_____	_____	_____	_____

Service Technician: Mike Johnson

Form 1 03/31/00